



Product Service

CERTIFICATE

No. Z1 12 09 44229 008

Holder of Certificate: **Invacare AG**
 Benkenstrasse 260
 4108 Witterswil
 SWITZERLAND

Certification Mark:



Product: **Standard wheelchairs**
Manual wheelchairs

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713008198



Date, 2012-09-14  (Torsten Zimmer)

Page 1 of 2



CERTIFICATE
No. Z1 12 09 44229 008

Model(s): **Advance**

Parameters:

total length incl. footrests:	766 mm (90° Frame), 860 mm (75° Frame)
width:	510 - 670 mm
height:	780 - 875 mm
seat width:	340, 360, 380, 400, 420, 440, 460, 480 and 500 mm
seat depth:	360, 380, 400, 420, 440, 460, 480, 500 and 520 mm
seat height in the front:	450 - 520 mm
backrest height:	270 - 480 mm
max. user weight:	125 kg
brake system:	knee lever brake (user brake)
back wheels:	
- size	22 x 1", 24 x 1", 25 x 1" and 26 x 1"
- tyres	air filled / puncture proof
- adjustable	yes
- fixation	quick release
front wheels:	
- size	Ø 100 x 32 mm and Ø 125 x 32 mm
- tyres	puncture proof
- adjustable	yes
- fixation	rigid
frame material:	aluminium
armrest:	not applicable
legrest:	height adjustable
anti tip device:	available

Tested according to: DIN EN 12183:2009

Production Facility(ies): 44229