

S8464

DISTRIBUTION

Invacare.:
Millbrook P.G.:

H. Thomas
Contract File

(1 Copy)
(Originals)

TEST DESCRIPTION

Test No.: S8464
Test Date: 17th Nov 2003
Test Type: ISO 7176/19
Wheelchair: Spectra plus with Seat Riser
Orientation: Forward Facing
Seat Restraint : Koller
Occupant: 50th %ile H II

CONTENTS

Graphical Results	Appendix A
Transducer Calibration Report	
Still Photographs	Appendix B
Film Analysis	Appendix C
Test Results	Appendix D

Prepared by: _____
A.de Boisgelin - Engineer

Date : _____

Approved by: _____
K. Forinton - Engineer

Date : _____

Test Results

ISO/FDIS 7176/19
30th August 2001

Section 5.2 - Dynamic Performance Requirements

Wheelchair: Occupant:	Spectra plus with Seat Riser Hybrid II Percentile (75kg)	RESULTS
5.2.1	During the Test	
5.3.1 a).	Horizontal ATD and wheelchair excursion limits as per limits shown in Table 3:-	
	Was the horizontal movement of the test wheelchair P- Point (X wc) less than 200 mm. (± 5 mm)	YES 86mm
	Was the horizontal movement of the dummy knee (X knee) less than 375 mm. (± 5 mm)	YES 201mm
	Was the horizontal movement of the dummy head (X head) less than 650 mm. (± 5 mm)	YES 350mm
	b). Was the ratio X knee/X wc >1.1	YES 2.34
	c). Not Measured	
	d). Did the batteries of powered wheelchairs, or their surrogate parts:-	
	i). move outside of the wheelchair footprint	NO
	ii). move into the wheelchair user's space	NO
5.2.2	Post Test	
5.3.2 a).	Did the wheelchair remain upright on the test platform Did the ATD remain in a seated posture in the test wheelchair with a torso angle less than 45°	YES
	b). Did the wheelchair securement points show visible signs of material failure	NO
	c). Did any components of a mass greater than 100g become detached from the wheelchair	NO See note 1
	d). Did any occupant contactable components fragment or separate with an edge of less than 2mm	NO
	e). Did any primary load carrying components of the wheelchair show any visible signs of failure	NO
	f). Did any 'tilt in space' locking mechanisms show signs of failure	NO
	g). Was the ATD released without the use of tools	YES
	f). Was the wheelchair released from the restraint system without the use of tools	YES
	h). Was the average decrease of H-Point height relative to the wheelchair platform less than 20% of the pre-test height.	YES 6%
	Has the wheelchair satisfied the Dynamic Test requirements of ISO/FDIS 7176/19 of 31/August/2003	YES

Note1: The battery cover detached but its movement was restrained by the lead.