



Product Service

# CERTIFICATE

No. Z1 12 10 42097 052

**Holder of Certificate:** Invacare Rea AB  
**Org. No. 556098-9211**  
 Växjövägen 303  
 343 71 Diö  
 SWEDEN

**Certification Mark:**



**Product:** Walking aids  
 Rollator

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 713005781



**Date,** 2012-10-31  (Torsten Zimmer)

Page 1 of 2



**CERTIFICATE**  
**No. Z1 12 10 42097 052**

**Model(s): Jazz 510 and Jazz 610**

**Parameters:**

max. user weight:	150 kg
max. load for basket:	7 kg
max. length:	690 mm
max. width:	595 mm
rollator height:	from 650 to 800 mm, from 750 to 1000 mm
turning width:	860 mm
folded dimensions (heightxwidthxlength):	220 x 690 x 800 mm
handgrip length:	115 mm
handgrip width:	38 mm
brake grip distance:	75 mm
wheel width:	30 mm
	(acc. to EN ISO 11199-2-2, 3.20)
wheel diameter:	Ø 200 mm
seat height:	510 and 610 mm
rollator weight: (without accessories)	7,5 and 7,8 kg
material:	
frame:	aluminum
push handles:	TPE
brake handles:	PA 6
seat plate:	PA 6 / PE cell foam
accessories:	basket

**Tested according to:** EN 1985:1998  
 EN ISO 11199-2:2005

**Production Facility(ies):** 33415